

Research Space

Journal article

A portable elbow exoskeleton for three stages of rehabilitation

Manna, S. and Dubey, V. N.

© ASME

CC-BY

A Portable Elbow Exoskeleton for Three Stages of Rehabilitation

Soumya K Manna, Venketesh N. Dubey

Faculty of Science and Technology, Bournemouth University

Talbot Campus, Poole, BH12 5BB, United Kingdom

Abstract: Patients suffering from stroke need to undergo a standard and intensive rehabilitation therapy. The rehabilitation training consists of three sequential stages; the first stage is controlled joint movement under external actuator, the second stage deals with supporting the movements by providing assistive force and the last stage provides variety and difficulty to exercises. Most of the exoskeletons developed so far for rehabilitation are restricted to a particular type of activity. Although a few exoskeletons incorporate different modes of rehabilitation, those are software controlled requiring sensory data acquisition and complex control architecture. To bridge this gap, a portable elbow exoskeleton has been developed for delivering three stages of rehabilitation in a single structure without affecting the range of motion and safety features. Use of electric motor and springs have been arranged in the actuation mechanism to minimise the energy consumption. The developed exoskeleton enhances torque to weight ratio compared to existing models and all three modes of rehabilitation have been controlled using a single motor.

Keywords: Exoskeleton, Rehabilitation, Portable, Stiffness control, Gravity compensation, Stroke.

1. Introduction

Present statistics shows that there are about 33 million stroke survivors worldwide [1]. The annual health and social costs of caring for disabled stroke patients are estimated to be in excess of £5 billion in the UK alone [2]. To reduce the burden of manual therapy which includes unavailability of sufficient number of caregivers and intensity of exercises, exoskeleton based rehabilitation has become a promising alternative [3]. However, the majority of developed exoskeletons can provide only a specific type of exercise [4]. A few standard rehabilitation processes are followed from acute stage to full recovery stage after stroke [5], [6], [7]. All these stages are used to regain the controlled muscle movement by reducing spasticity and involuntary movement. After analysing the function and treatment procedure of each stage, these can be categorized into three distinct stages as shown in Fig. 1. These are actuator based joint control, supportive force and resistive force.




Figure 1. Three phases of rehabilitation for post-stroke patients

If exercise is performed in transverse plane using exoskeleton, patients only need to overcome the frictional force of the exoskeleton structure. If the same exercise is performed in sagittal plane, the

exoskeleton requires higher joint torque to carry out the load against gravity; size of the actuator as well as the actuation system needs to be effective to obtain the required torque. The range of movement is improved if exercises were performed in gravity compensated training environment [8]. Postural stability is achieved by active holding of the body segment against external force [9].

Even after three decades of research, no standard solution has been presented for the design of exoskeleton to provide the best rehabilitation therapy [10]. Most of the exoskeletons have focused on design aspect which includes portability and user-friendliness but failing on providing the standard rehabilitation training. Two essential factors can be considered for the design of exoskeletons; one is its mechanical design and other is the rehabilitation therapy.

The quality of rehabilitation therapy using exoskeleton can be improved by introducing three stages of rehabilitation in a single structure which can possibly be accomplished in two ways; one is the hardware-based solution and the other is software approach. In software approach [11], actuator used in the exoskeleton can provide three stages of therapy using adaptive control algorithms. In order to execute patient-oriented exercises using external actuators, electric motors are normally placed at the joint of most exoskeletons [12]. Here different bio-sensors attached to user send signals to the control system about the patient's intention and the motors provide assistive or resistive torque to the affected joint as per the signal received from biosensors (EMG, EEG). Because of the reliance on biosignals, those systems are inoperable without sensors. Movement based on EMG data extraction from stroke patients is difficult because of abnormal EMG-torque relationship [13]. The adaptive control algorithm used in exoskeleton results in constant draining of energy for controlling the variable joint torque and active range of motion. Also, the adaptive control system always takes over the control by making patients inactive which indirectly reduces their activities during training [14], therefore diminishes the rehabilitation effectiveness. Joint-based actuation system also requires higher torque compared to the designs where joint is remotely controlled. To carry out the exercises with higher load, size and weight of the motor are also increased [15] and so is the cost. As a result, most of the electric motor controlled exoskeletons are ground-based system [16]. In the former case, human joint is always under motor control which might not be ideal from safety point of view. If the joint moves beyond the anatomical range, it may cause injury. Looking at these limitations, a hardware-based solution may provide viable option for user acceptance.

Integration of multistage rehabilitation can be achieved using active and passive components in the mechanism which can reduce the complexity of the control system. For example, a system can use an electromagnetic clutch/brake for shifting from one rehabilitation mode to another though it will drain energy and create unwanted noise during switching. Passive actuation systems use elastic elements such as spring or rubber band which can provide the required joint torque for reducing the gravity force during elbow movement. Such spring-based exoskeletons [17],[18] do not need any energy source to actuate but these systems can only provide assistive force to users. The back-drivable motor in combination with a series elastic actuator [19] is also able to provide both types of rehabilitation, however if the back-drivability is too low, the gearbox can be damaged due to sudden external force. Compliant mechanism [20] can provide variable stiffness to the joint however, it can only generate resistive force.

Hence it is quite challenging to integrate all types of exercise in a single exoskeleton because exercises involved in the three stages after stroke are totally different in nature. In acute stage, patients require fixed contact to human arm because they have no power left to move their arm but in rest of the recovery stages, exoskeleton needs a compliant contact to allow them to carry out the exercises themselves. The aim of this paper is to create an innovative joint mechanism for the exoskeleton to achieve both these properties without any extra burden to system or risk to the users. The novelty of the developed exoskeleton can be described as:

- The developed exoskeleton delivers all modes of exercises (external force, assistive and resistive) required for three stages of rehabilitation in a single structure.
- The exoskeleton mechanism generates variable assistive as well as resistive force without using any complex control algorithms.

- Electric motor is used to control the joint whereas in rest of the two modes, joint motions are supported by stiffness of the springs for providing assistive or resistive force.
- Spring stiffness is used for the switching mechanism to shift between rehabilitation modes, therefore, no brakes or clutches are required making it an energy efficient mechanism.
- The switching mechanism supports safety of the users mechanically since joint control is transferred from motor to user when the elbow joint goes beyond the permissible limit.
- To achieve all this a single motor has been used in the whole exoskeleton design.

2. Design description

The exoskeleton has been conceptualized as a mechanism where the whole operating region is divided into three sub-regions to provide specific exercises as shown in Fig. 2. All these regions are interconnected and will appear one after another mechanically. The type of exercise generated by exoskeleton is aligned with the post-stroke recovery stages. The exoskeleton utilizes the motor torque in acute phase when users do not have enough strength and provides spring energy during self-movement for generating assistive and resistive force. A couple of springs (compression and torsional) have been used in the exoskeleton for switching between different regions. The schematic diagram of the exoskeleton and its 3D model are shown in Fig. 3 (a and b).






Figure 2. Operating region of the exoskeleton



(a) Schematic diagram



(b) 3D model

Figure 3. Exoskeleton Design

(1) Baseplate	(7) Nut slider	(13) Connecting link
(2) Motor	(8) Concentric slider	(14) Universal joint
(3) Gear	(9) Elbow joint	(15) Claw-type jaws
(4) Solid rods	(10) Revolute joint	(16) Rectangular slider
(5) Slider for variable stiffness	(11) Compression spring	(17) Connected plates
(6) Leadscrew	(12) Forearm supporting link	(18) Small cylindrical rod

The relationship between the distance covered by the nut slider (x) and the motor rotations (n, θ) is given by

$$x = \frac{n\theta L}{2\pi N} \quad (1)$$

Where, n - Number of turns of the motor

θ - Angle made by the motor

L - Lead of the screw

N - Gear ratio for transferring the motion to the leadscrew (1.5:1)

Mode of rehabilitation: $0 \leq x \leq x_1$ = Electric motor based joint control (0 to 0.18 m)

$x_1 < x \leq x_2$ = Spring based assistive force (0.181 to 0.20 m)

$x_2 < x \leq x_3$ = Spring based resistive force (0.201 to 0.24 m)

Where x_1, x_2, x_3 are the switching positions.

2.1 Electric motor based joint control (first region)

In the first region, the electric motor controls the joint movement without any active participation from the user. The actuation system has been designed based on a leadscrew in combination with a slider-crank mechanism (Fig. 4). Motion from the motor is transferred to the leadscrew through reduction gears. Slider-crank mechanism converts the linear motion of leadscrew into elbow joint rotation. The leadscrew and slider crank mechanism are not directly coupled to each other. There are two loads to overcome; one of which acts as a nut that translates in both directions on the screw and the other slider moves over the leadscrew concentrically without being placed on the leadscrew. In the first region ($0 < x \leq x_1$), a spring (S_6) actuates the locking mechanism that keeps both sliders in contact as a single unit until the elbow joint rotates to its maximum anatomical limit (0-130°).





Figure 4. Electric motor based joint control

2.2 Switching from the electric motor control (first region) to assistive force (Second region)

In the locking mechanism, two claw-type jaws are connected to the nut slider in the form of a four-bar mechanism (Fig. 5(a), Locked condition). The locking condition remains enforced until the two compression springs S_5 and S_6 clash with each other due to the backward movement of the nut slider. As soon as the nut slider crosses the switching position ($x > x_l$), switching takes place. Because of the higher stiffness, the force exerted by S_5 is greater than S_6 , thus a small displacement of S_5 causes a large displacement in S_6 . As a result, S_6 will be compressed by the resultant force and both jaws will rotate about a fixed point to free those sliders (see the unlocked condition, Fig. 5(b)). However, forward movement of the nut slider beyond the switching point will restore the locking mechanism again. The ratio of the stiffness of S_5 and S_6 has been determined in a way that the switching region becomes as small as possible.



(a) Locked condition (b) Unlocked condition

Figure 5. Switching from motor based control to spring assisted force

After opening of the lock, the nut slider and concentric slider are detached from each other and the joint rotation is not under electric motor control. In this phase, patients are free to control their own movements and S_2 provides assistive force to help rotate the elbow joint (Fig. 6). The joint is torque balanced at every configuration due to the spring force and only a small effort is required from patients for lifting up any load or the forearm loads against gravity. Higher assistive force reduces the effort of users to reach a full joint rotation during flexion. The same assistive force opposes the forearm freefall during extension. In this way, the assistive force balances the arm weight and slows down the joint movement to achieve full extension. The assistive force during rehabilitation should be adaptable for different load under gravity. The structural part of the exoskeleton to provide variable gravity compensation consists of two torsional springs (S_7 and S_8), one compression spring (S_1), one small cylindrical rod (CR_1), one small rectangular slider (SL_1) and two rectangular plates (RP_1 and RP_2). SL_1 is concentric to CR_1 which is attached to the base plate. The range of spring force provided by S_2 can be amplified by changing the span of displacement. RP_1 and RP_2 are connected to SL_1 using S_7 and S_8 on both sides in such a way that these plates can rotate about the axis of these

torsional springs (see the magnified view below). RP_1 and RP_2 have been used to maintain the force balancing condition during rehabilitation to provide a constant supply of assistive force. CR_1 has a rectangular channel to provide a guiding path to SL_1 . The guiding path has two mechanical restrictions for controlling the movement of SL_1 within a particular range. This is the region where different spring force can be generated. The role of S_1 is to restore the whole setup to its original position once released.




Figure 6. Spring configurations during assistive force mode

At the initial condition of self-initiated joint movement, the front end of S_2 is fixed which allows a fixed range of spring force. To increase the spring force dynamically, the front-end of S_2 is shifted backward near the baseplate; the extended part of nut slider has been utilized for this purpose. The backward movement of the nut slider in this region pushes RP_1 and RP_2 connected to S_7 and S_8 . The stiffness of S_7 and S_8 is high enough to be deflected by a small force, as a result, the whole arrangement connected to SL_1 will move backward along with the nut slider. Due to the torsional stiffness, S_7 and S_8 create an opposing torque which is equalized by the reaction force from the nut slider during the movement. The second mechanical restriction on the guiding path does not allow SL_1 to move further in the backward direction. This is the position where the mechanism can develop maximum assistive force at the joint using S_2 . Therefore, further pressure from the nut slider will put S_7 and S_8 beyond their limit and RP_1 and RP_2 are deflected to come out from the range of nut slider. Because of the stiffness property of S_1 , SL_1 will come to its initial position with all its arrangement.

2.3 Resistance based rehabilitation (Third region)

In this region, the exoskeleton provides a resistive force to the joint to restrict its motion which is achieved by changing the elbow joint stiffness. Two pairs of extension springs (S_3 and S_4) are connected in parallel to the end of elbow joint on both sides and can slide on two solid parallel rods. Backward movement of the nut slider beyond the region ($x > x_2$) will stretch both S_3 and S_4 resulting in higher joint stiffness (Fig. 7). Two linear springs (S_3 and S_4) of different stiffness have been used in this design.




Figure 7. Spring configuration in resistive force mode for variable joint stiffness

The mechanism of the exoskeleton has been improved from other designs to get more flexibility such that a universal joint is used to replace the normal revolute joint for elbow movement so that the forearm is not fixed during flexion and extension. The universal joint provides a slight lateral movement of ($\pm 5^\circ$). Out of the two degrees of freedom possessed by the universal joint, active one is responsible for flexion-extension of the elbow whereas the passive joint supports flexibility in the transverse plane during joint rotation. Another universal joint is used at the junction between the lead screw slider and the connecting link to support the joint flexibility. The user's forearm is attached to the exoskeleton's forearm using cuff and straps. The forearm supporting link has discrete holes to fit different arm lengths. To maintain the alignment of the centre of rotation between the exoskeleton and user, the forearm has a passive translational joint with a compression spring whose length is varied to match the forearm length of the user as shown in Fig. 3.

3. Torque analysis for motor control

In the first region, the motor torque required to actuate the elbow joint is equivalent to the torque needed to overcome the frictional force created between the lead screw and the nut slider. Due to the slider-crank mechanism, the elbow joint is actuated by pulling the connecting link (Fig. 8).




Figure 8. Mechanism analysis during electric motor control

The required force (T) for lifting up the weight of the forearm is given by

$$T = \frac{Mg L_1 \cos \beta}{r \cos(\alpha - \beta)} \quad (2)$$

Where, M - Mass of the forearm and the supporting link

L_1 - Distance from the elbow joint to the centre of gravity of forearm and the supporting link

β - Elbow joint angle

α - Angle made by the connecting link and nut slider

r - Length of the crank

g - Acceleration due to gravity

From the model of the lead screw (Fig. 8), it can be shown, $\tan \delta = \frac{p}{\pi d_1}$ (3)

Where p - Pitch of the lead screw, d_1 - Diameter of the lead screw and δ - Lead angle of lead screw

232 P is effort applied to the screw to lift the load. Taking the force equilibrium (Fig. 8),
 233 We have, $P \cos \delta = W_1 \cos \delta + W \sin \delta + F$ [Where $T \cos \alpha = W_1$ and $T \sin \alpha = W$] (4)

234
 235 Frictional force (F) during motion is
 236 $F = \mu R_N = \mu(W_1 \sin \delta - W \cos \delta - P \sin \delta)$ [where μ = coefficient of friction] (5)

237 After substituting the value of F and $\mu = \tan \varphi$ (φ is friction angle)] in Eq. (4), we get
 238 $P = W_1 + W \tan(\delta - \varphi)$ (6)

239
 240 Putting the value of W and W_1 , torque (τ) required for overcoming the friction of the leadscrew is
 241 $\tau = P \times \frac{d_1}{2} = \frac{T(\cos \alpha + \sin \alpha \tan(\delta - \varphi))d_1}{2}$ (7)

242 Putting the value of T from Eq. (2), the final equation for required motor torque (τ) of the
 243 exoskeleton is $\tau = \frac{MgL_1 \cos \beta (\cos \alpha + \sin \alpha \tan(\delta - \varphi))d_1}{2r \cos(\alpha - \beta)}$ (8)

244 The relation between α and β can be derived from Fig. 8 as $\alpha = \cos^{-1} \left(\frac{d-r \sin \beta}{l} \right)$ (9)

245 If motor is placed directly on the joint, the required motor torque is $\tau' = MgL_1 \cos \beta$ (10)

246 Fig. 9 shows the required motor torque in two configurations which is significantly reduced for the
 247 proposed exoskeleton.




Figure 9. Comparison of the motor torque in two configurations

250 4. Selection of springs and their stiffness calculation

251 Since the linear springs (S_1, S_2, S_3, S_4, S_5 and S_6) and torsional springs (S_7 and S_8) are used in the
 252 design either for providing the spring force or switching from one stage of rehabilitation to another,
 253 therefore, the stiffness of all springs must be determined for the exoskeleton design.

254 4.1 Stiffness of S_1

255 The function of S_1 is to restore the position of the front-end of S_2 at the end of the assistive force
 256 region, therefore, the stiffness of S_1 needs to be high enough to overcome the frictional force
 257 between SL_1 (along with the all other components connected to it) and CR_1 , see Fig. 10.





Figure 10. S_1 in fully compressed state

260 Therefore, $K_1 x_{s1} > \mu m g$ (11)
 261 Where K_1 - Stiffness of S_1
 262 x_{s1} - Displacement covered by S_1 in fully compressed position
 263 μ - Coefficient of friction between SL_1 and CR_1
 264 m - Weight of the assembly connected to SL_1
 265 g - Acceleration due to gravity
 266 Based on the frictional property of SL_1 and CR_1 and mass of SL_1 , frictional force can be determined,
 267 from where K_1 can be estimated. S_1 produces maximum force when it is fully compressed ($x = x_2$) by
 268 the nut slider.

269 4.2 Stiffness of S_2

270 Spring force of S_2 is mainly responsible for assisting the elbow movement in the second region (Fig.
 271 11). In the exoskeleton, S_2 will be extended for sharing the required torque used to rotate the joint
 272 against gravity.



273 Figure 11. Force balancing in the assistive force mode (second region)
 274

275 As linear bearings are used at the sliding contact between the concentric slider and leadscrew, the
 276 frictional force during motion is considered negligible compared to the elbow actuation force and is
 277 not taken into account.

278 The assistive force provided by S_2 is $f_{s2} = K_2(x_{s2} - x'_{s2})$ (12)

279 Where K_2 - Stiffness of S_2 , x_{s2} - Displacement of S_2 , x'_{s2} - Free length of S_2 .

280 Pulling force (T) along the connecting link is same as it is shown during the electric motor control.
 281 Therefore, the value of T is taken from Eq. (2). The only difference is that S_2 is taking the load
 282 instead of the motor.

283 Therefore, by equilibrating forces in Fig. 11, the stiffness of S_2 becomes

$$284 K_2 = \frac{MgL_1 \cos \beta \sin \alpha}{r \cos(\alpha - \beta)(x_{s2} - x'_{s2})} \quad (13)$$

285 The displacement range of S_2 can be increased by pushing the nut slider backward towards the
 286 baseplate, thus providing more assistive force.

287

288 4.3 Stiffness of S_3 and S_4

289 The stiffness of S_3 and S_4 are used for changing the joint stiffness providing the resistive force (Fig.
 290 12). The elbow joint stiffness is dependent on three springs (S_2 , S_3 and S_4). However, the spring
 291 parameters of S_2 are constant during the resistive force control, only the displacement of S_3 and S_4
 292 is changed to create a variable joint stiffness at the elbow joint.




Figure 12. Elbow exoskeleton during variable joint stiffness control

The component of the spring force exerted by S_2 about the point C is given by,

$$f'_{s2} = \frac{K_2(x_{s2} - x'_{s2}) \cos(\alpha - \beta)}{\sin \alpha} \quad (14)$$

Spring force exerted by S_3 ,

$$f_{s3} = K_3(x - r_1\beta - x'_{s3}) \quad [\text{Where } K_3 - \text{Stiffness of } S_3, x'_{s3} - \text{Free length of } S_3] \quad (15)$$

Spring force exerted by S_4 ,

$$f_{s4} = K_4(x + r_1\beta - x'_{s4}) \quad [\text{Where } K_4 - \text{Stiffness of } S_4, x'_{s4} - \text{Free length of } S_4] \quad (16)$$

Therefore the joint stiffness of the elbow joint is the torsional stiffness K' which is given by

$$K' = \frac{\tau}{\beta} = \frac{r_1(f_{s4} - f_{s3})}{\beta} - \frac{rf'_{s2}}{\beta} \quad (17)$$

Where r_1 - Radius of the pulley connected at the elbow joint

Two pairs of S_3 and S_4 are connected in this mechanism; therefore, the force exerted by both springs will be doubled. Substituting the value of f_{s2} , f_{s3} and f_{s4} in Eq. (17), the elbow joint stiffness is given by

$$K' = \frac{2r_1\{(K_4 - K_3)x - (K_4x'_{s4} - K_3x'_{s3})\} - \frac{K_2r(x_{s2} - x'_{s2}) \cos(\alpha - \beta)}{\sin \alpha}}{\beta} + 2r_1^2(K_4 + K_3) \quad (18)$$

The joint stiffness variation is shown in Fig. 13.




Figure 13. Elbow joint stiffness variation for different position of the nut slider

4.4 Stiffness of S_5 and S_6

The stiffness of both compression springs (S_5 and S_6) used for locking operation is equally important in switching operation between first and second region. The ratio of the stiffness of S_5 and S_6 depends on the construction parameters of the locking mechanism (Fig. 14).




Figure 14. Force balancing during the unlocked condition

In the locking mechanism, the right-handed jaw KMPO is shown. The upper-end position of the jaw O should be outside of the region covered by these sliders (shown in dotted line OM). It is clear that these two jaws need to rotate a minimum angle φ about point M to unlock the concentric slider from the locking range. Here, ΔOMP is a right-angled triangle. Therefore, the required angle φ for unlocking concentric slider from the jaw is

$$\varphi = \tan^{-1} \frac{e}{d} \quad (19)$$

Length of the solid links used in this mechanism are a , b , c , d , e and f . $\angle MLI$ and $\angle KMP$ are also a part of the structure. As the values of e and d are constant, the value of φ is defined for the unlocking condition. To achieve the angle, S_6 needs to move a particular distance which can be derived by the geometrical parameters.

The locking mechanism can function successfully if it satisfies the following condition, $K_5 \gg K_6$ [Where K_5 - Stiffness of S_5 , K_6 - Stiffness of S_6] which means, $x_{s5} \ll x_{s6}$

At the time of opening the lock, both springs will be in equilibrium which means force exerted by S_5 and S_6 will be the same at that position.

Therefore, $f_5 = f_6$

i.e. $K_5 x_{s5} = K_6 x_{s6}$

$$K_5 = \frac{K_6 x_{s6}}{x_{s5}} \quad (20)$$

$$\text{From Fig. 14, } \cos \angle JIK = \frac{x_{s6}^2 + KI^2 - a^2}{2x_{s6}KI}$$

$$\cos(90^\circ - \angle KIL) = \frac{x_{s6}^2 + KI^2 - a^2}{2x_{s6}KI} \quad (21)$$

Using trigonometric relation, it can be found that

$$KI^2 = f^2 + A_1 - 2f\sqrt{A_1} \cos\left(\delta - \cos^{-1}\left(\frac{B_1}{\sqrt{A_1}}\right)\right) \quad (22)$$

$$\text{and } \angle KIL = \cos^{-1}\left(\frac{f - \sqrt{A_1} \cos\left(\delta - \cos^{-1}\left(\frac{B_1}{\sqrt{A_1}}\right)\right)}{f^2 + A_1 - 2f\sqrt{A_1} \cos\left(\delta - \cos^{-1}\left(\frac{B_1}{\sqrt{A_1}}\right)\right)}\right) \quad (23)$$

$$\text{Where } A_1 = b^2 + c^2 + 2bccos\left(\delta + \tan^{-1} \frac{e}{d}\right)$$

$$B_1 = c + bcos\left(\delta + \tan^{-1} \frac{e}{d}\right)$$

After calculating the value of x_{s6} , x_{s5} and putting in Eq. (20), we get relationship between K_5 and K_6 .

The spring (S_6) associated with the lock will experience a higher and opposite force from S_5 . After opening of the lock, S_6 cannot be compressed further due to the mechanical constraint thus exhibiting a constant force for the rest of the motion. However, due to the movement of nut slider in backward direction, S_5 will be further compressed with a higher spring force which helps to maintain the unlocked condition during rest of the range as shown Fig. 15.




Figure 15. Force generation in two springs of the locking mechanism

4.5 Stiffness of S_7 and S_8

Fig. 16 shows the force balancing of the mechanism during the final stage of the assistive mode where both torsional springs S_7 and S_8 are at their maximum deflected position. The nut slider generates an equal and opposite force against two torsional springs (S_7 and S_8) and balances the forces generated by S_1 and S_2 . Both S_7 and S_8 have equal stiffness as they are structurally the same.




Figure 16. Force balancing of the mechanism at the final stage of assistive force mode

From Fig. 16, it can be derived that,

$$\lambda = \cos^{-1}\left(1 - \frac{a'}{r_2}\right) \quad (24)$$

Where a' - Width of the extension part of the nut slider, r_2 - Length of the plates (RP₁ and RP₂)

$$\text{Force produced by } S_1, f_{s1} = K_1 x_{s1} \quad [\text{Where } K_1 - \text{Stiffness of } S_1] \quad (25)$$

361 Force produced by S_2 , $f_{s2} = K_2(x_{s2} + x_{s1} - x'_{s2})$ (26)

362 For this mechanism, $K_7 = K_8$ [Where K_7 - Stiffness of S_7 and K_8 - Stiffness of S_8]

363 From Fig. 16, $\tau' = K_7\lambda$

364 Where λ - Angle made RP_1 and RP_2 at maximum deflected position

365 τ' - Torque created by S_7 and S_8

366 $F'r_2 \cos \lambda = K_7\lambda$ Where F' - Reaction force by S_7 and S_8

367 Therefore, $F' = \frac{K_7\lambda}{r_2 \cos \lambda}$ (27)

368 As S_7 and S_8 maintain the force in equilibrium,

369 $2F' = f_{s1} + f_{s2}$ (28)

370 Putting the value of f_{s1} (from Eq. (25)), f_{s2} (from Eq. (26)) and F' (from Eq. (27)) in Eq. (28), we

371 have $K_7 = \frac{r_2 \cos \lambda (K_1 x_{s1} + K_2 (x_{s2} + x_{s1} - x'_{s2}))}{2\lambda}$ (29)

372 After putting the value of λ (taken from Eq. (24)), the value of K_7 will be,

373 $K_7 = \frac{(r_2 - a') (K_1 x_{s1} + K_2 (x_{s2} + x_{s1} - x'_{s2}))}{2 \cos^{-1}(1 - \frac{a'}{r_2})}$ (30)

374 Based on the above design considerations, a functional prototype has been developed to establish the
 375 working principle of the exoskeleton. All mechanical components have been manufactured using 3D
 376 printer. The prototype of the elbow exoskeleton is shown in Fig. 17 along with its specifications. All
 377 the sliding contacts have been developed with bearing to reduce the frictional loss during motion.
 378 The prototype performs as per the requirements of the three stages of rehabilitation.
 379



Material	ABS (Acrylonitrile butadiene styrene)
Upper arm dimension	0.35 m x 0.15 m x 0.17 m
Forearm dimension	0.40 m x 0.02 m x 0.06 m
Weight	1.8 kg
Motor specification	Model-DFROBOT ZYTD520
	Operating voltage-12V DC
	RPM-50 RPM
	Maximum torque-4.6 Nm
	Weight-0.210 kg
	Power - 5 W
Gear material	Nylon-101
Spring material	ASTM A228
Linear Ball Bearing	Model- Bosch Rexroth Linear Ball Bearing R060204010
Axial Ball bearing	HCH 62022

Figure 17. Prototype of the elbow exoskeleton with specifications

383 5. Conclusions

384 An innovative mechanism of the elbow exoskeleton has been developed which can accommodate
 385 three modes of rehabilitation for different stages after stroke. In this design, we have attempted to
 386 achieve the multistage post-stroke rehabilitation at mechanical level so that the device can be fine-
 387 tuned to user's requirements. Full design details of the elbow exoskeleton have been presented
 388 together with parametric relations for component selection. These design parameters can be tailored
 389 to suit any user specific requirements. A prototype device has been developed to prove the principle.
 390 For most exoskeletons, the motor torque is varied depending on the dynamics of the model and
 391 patient's requirement whereas in this exoskeleton the position of the nut-slider can produce different
 392 exercise modes either under motor control or in assistive or resistive modes. The mechanism can
 393 change the amount of assistive and resistive force by simply changing the position of the slider. Such
 394 arrangement in a single structure offers flexibility to patients to select a particular type of exercise.

During the assistive and resistive modes only the spring force is used without engaging any active actuator therefore the energy source is only used during the motor operation. The switching mechanism safeguards users by restricting the reachable joint angle to the anatomical limit during motor control mode. If the nut slider goes up to the end of first rehabilitation region due to motor rotation, the lock will be activated till the joint takes its maximum anatomical limit. Position of the nut slider beyond the switching point will automatically open the lock, releasing the joint control from motor and transfer it to the user, therefore providing safety and functionality at the same time.

References

- [1] Feigin, V. L., Forouzanfar, M. H., Krishnamurthi, R., Mensah, G. A., Connor, M., Bennett, D. A., Moran, A. E., Sacco, R. L., Anderson, L., and Truelsen, T., 2014, "Global and regional burden of stroke during 1990–2010: findings from the Global Burden of Disease Study 2010," *The Lancet*, 383(9913), pp. 245-255.
- [2] Saka, Ö., McGuire, A., and Wolfe, C., 2009, "Cost of stroke in the United Kingdom," *Age and ageing*, 38(1), pp. 27-32.
- [3] Lo, A. C., Guarino, P. D., Richards, L. G., Haselkorn, J. K., Wittenberg, G. F., Federman, D. G., Ringer, R. J., Wagner, T. H., Krebs, H. I., and Volpe, B. T., 2010, "Robot-assisted therapy for long-term upper-limb impairment after stroke," *New England Journal of Medicine*, 362(19), pp. 1772-1783.
- [4] Manna, S. K., and Dubey, V. N., 2016, "Upper arm exoskeleton –what specifications will meet users' acceptability?," *Robotics: New Research* D. G. Fisher, ed., Nova Science Publisher, pp. 123-169.
- [5] Proietti, T., Crocher, V., Roby-Brami, A., and Jarrassé, N., 2016, "Upper-limb robotic exoskeletons for neurorehabilitation: a review on control strategies," *IEEE Reviews in Biomedical Engineering*, 9, pp. 4-14.
- [6] Pineda-Rico, Z., de Lucio, J. A. S., Martinez Lopez, F. J., and Cruz, P., 2016, "2121. Design of an exoskeleton for upper limb robot-assisted rehabilitation based on co-simulation," *Journal of Vibroengineering*, 18(5).
- [7] Chonnaparamutt, W., and Supsi, W., 2016, "SEFRE: Semiexoskeleton Rehabilitation System," *Applied Bionics and Biomechanics*, 2016.
- [8] Beer, R. F., Naujokas, C., Bachrach, B., and Mayhew, D., "Development and evaluation of a gravity compensated training environment for robotic rehabilitation of post-stroke reaching," 2008, *Proc. 2nd IEEE RAS & EMBS International Conference on Biomedical Robotics and Biomechatronics*, IEEE, pp. 205-210.
- [9] Kolar, P., 2014, *Clinical rehabilitation*, Alena Kobesová.
- [10] Jarrassé, N., Proietti, T., Crocher, V., Robertson, J., Sahbani, A., Morel, G., and Roby-Brami, A., 2014, "Robotic exoskeletons: a perspective for the rehabilitation of arm coordination in stroke patients," *Frontiers in Human Neuroscience*, 8, p. 947.
- [11] Wolbrecht, E. T., Chan, V., Reinkensmeyer, D. J., and Bobrow, J. E., 2008, "Optimizing compliant, model-based robotic assistance to promote neurorehabilitation," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, 16(3), pp. 286-297.
- [12] Maciejasz, P., Eschweiler, J., Gerlach-Hahn, K., Jansen-Troy, A., and Leonhardt, S., 2014, "A survey on robotic devices for upper limb rehabilitation," *Journal of Neuroengineering and Rehabilitation*, 11(1), p. 3.
- [13] Bhadane, M., Liu, J., Rymer, W. Z., Zhou, P., and Li, S., 2016, "Re-evaluation of EMG-torque relation in chronic stroke using linear electrode array EMG recordings," *Scientific Reports*, 6, p. 28957.
- [14] Marchal-Crespo, L., and Reinkensmeyer, D. J., 2009, "Review of control strategies for robotic movement training after neurologic injury," *Journal of Neuroengineering and Rehabilitation*, 6(1), p. 20.
- [15] Marcheschi, S., Salsedo, F., Fontana, M., and Bergamasco, M., "Body extender: whole body exoskeleton for human power augmentation," *Proc. 2011 IEEE International Conference on Robotics and Automation (ICRA)*, IEEE, pp. 611-616.
- [16] Manna, S. K., and Dubey, V. N., 2018, "Comparative study of actuation systems for portable upper limb exoskeletons," *Medical Engineering & Physics*, 60, pp. 1-13.
- [17] Housman, S. J., Le, V., Rahman, T., Sanchez, R. J., and Reinkensmeyer, D. J., "Arm-training with T-WREX after chronic stroke: preliminary results of a randomized controlled trial," *Proc. IEEE 10th International Conference on Rehabilitation Robotics*, 2007, IEEE, pp. 562-568.
- [18] Sanchez, R., Reinkensmeyer, D., Shah, P., Liu, J., Rao, S., Smith, R., Cramer, S., Rahman, T., and Bobrow, J., "Monitoring functional arm movement for home-based therapy after stroke," *Proc. 26th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, 2004, IEEE, pp. 4787-4790.
- [19] Crea, S., Cempini, M., Moisè, M., Baldoni, A., Trigili, E., Marconi, D., Cortese, M., Giovacchini, F., Posteraro, F., and Vitiello, N., "A novel shoulder-elbow exoskeleton with series elastic actuators," *Proc. 2016 6th IEEE International Conference on Biomedical Robotics and Biomechatronics (BioRob)*, IEEE, pp. 1248-1253.
- [20] Van Ham, R., Sugar, T. G., Vanderborght, B., Hollander, K. W., and Lefeber, D., 2009, "Compliant actuator designs," *IEEE Robotics & Automation Magazine*, 16(3).